



**PLEASE CHECK FOR CHANGE INFORMATION  
AT THE REAR OF THIS MANUAL.**

**2 2 1 3  
OSCILLOSCOPE  
OPERATORS**

**I N S T R U C T I O N   M A N U A L**

**Tektronix, Inc.**  
**P.O. Box 500**  
**Beaverton, Oregon 97077**

070-3397-00  
Product Group 46

Serial Number \_\_\_\_\_

First Printing JUN 1981  
Revised AUG 1982

# CONTROLS, CONNECTORS, AND INDICATORS

The following descriptions are intended to familiarize the operator with the location, operation, and function of the instrument's controls, connectors, and indicators.

## POWER, DISPLAY, AND PROBE ADJUST

Refer to Figure 3 for location of items 1 through 7.

- ① **Internal Graticule**-Eliminates parallax viewing error between the trace and graticule lines. Rise-time amplitude and measurement points are indicated at the left edge of the graticule.
- ② **POWER Switch**-Turns instrument power on and off. Press in for ON; press again for OFF.
- ③ **AUTO FOCUS Control**-Adjusts display for optimum definition. Once set, the focus of the crt display will




Figure 3. Power, display, and probe adjust controls, connector, and indicator.

be maintained as changes occur in the intensity level of the trace.

- ④ **PROBE ADJ Connector**-Provides an approximately 0.5 V, negative going, square-wave voltage (at approximately 1 kHz) that permits the operator to compensate voltage probes and to check operation of the oscilloscope vertical system. It is not intended to verify the accuracy of the vertical gain or time-base calibration.
- ⑤ **BEAM FIND Switch**-When held in, compresses the display to within the graticule area and provides a visible viewing intensity to aid in locating off-screen displays.
- ⑥ **TRACE ROTATION Control**-Screwdriver control used to align the crt trace with the horizontal graticule lines.
- ⑦ **AUTO INTENSITY Control**-Adjusts brightness of the crt display. This control has no effect when the BEAM FIND switch is pressed in. Once the control is set, intensity is automatically maintained at approximately the same level between SEC/DIV switch settings from 0.5 ms per division to 0.05  $\mu$ s per division.

## VERTICAL

Refer to Figure 4 for location of items 8 through 16.

- ⑧ **SERIAL and Mod Slots**-The SERIAL slot is imprinted with the instrument's serial number. The Mod slot contains the option number that has been installed in the instrument.
- ⑨ **CH 1 OR X and CH 2 OR Y Connectors**-Provide for application of external signals to the inputs of the vertical deflection system or for an X-Y display. In the X-Y mode, the signal connected to the CH 1 OR X connector provides horizontal deflection, and the signal connected to the CH 2 OR Y connector provides vertical deflection.
- ⑩ **GND Connector**-Provides direct connection to instrument chassis ground.




Figure 4. Vertical controls and connectors.

- ⑪ **Input Coupling (AC-GND-DC) Switches**-Used to select the method of coupling input signals to the vertical deflection system.

**AC**-Input signal is capacitively coupled to the vertical amplifier. The dc component of the input signal is blocked. Low-frequency limit (-3 dB point) is approximately 10 Hz.

**GND**-The input of the vertical amplifier is grounded to provide a zero (ground) reference-voltage display (does not ground the input signal). This switch position allows precharging the input coupling capacitor.

**DC**-All frequency components of the input signal are coupled to the vertical deflection system.

- ⑫ **CH 1 VOLTS/DIV and CH 2 VOLTS/DIV Switches**-Used to select the vertical deflection factor in a 1-2-5 sequence. To obtain a calibrated deflection factor, the VOLTS/DIV variable control must be in detent.

**1X PROBE**-Indicates the deflection factor selected when using either a 1X probe or a coaxial cable.

**10X PROBE**-Indicates the deflection factor selected when using a 10X probe.

- ⑬ **VOLTS/DIV Variable Controls**-When rotated counter clockwise out of their detent positions, these controls provide continuously variable, uncalibrated deflection factors between the calibrated settings of the VOLTS/DIV switches. Extends maximum uncalibrated deflection factor to 25 volts per division with 1X probe (a range of at least 2.51).

- ⑭ **INVERT Switch**-Inverts the Channel 2 display when button is pressed in. Push button must be pressed in a second time to release it and regain a noninverted display.

- ⑮ **VERTICAL MODE Switches**-Two three-position switches are used to select the mode of operation for the vertical amplifier system.

**CH 1**-Selects only the Channel 1 input signal for display.

**BOTH**-Selects both Channel 1 and Channel 2 input signals for display. The BOTH position must be selected for either ADD, ALT, or CHOP operation.

**CH 2**-Selects only the Channel 2 input signal for display.

**ADD**-Displays the algebraic sum of the Channel 1 and Channel 2 input signals.

**ALT**-Alternately displays Channel 1 and Channel 2 input signals. The alternation occurs during retrace at the end of each sweep. This mode is useful for viewing both input signals at sweep speeds from 0.05  $\mu$ s per division to 0.2 ms per division.

**CHOP**-The display switches between the Channel 1 and Channel 2 input signals during the sweep. The switching rate is approximately 250 kHz. This mode is useful for viewing both Channel 1 and Channel 2 input signals at sweep speeds from 0.5 ms per division to 0.5 s per division.

- ⑯ **POSITION Controls**-Used to vertically position the display on the crt. When the SEC/DIV switch is set to X-Y, the Channel 2 POSITION control moves the display vertically (Y-axis), and the Horizontal POSITION control moves the display horizontally (X-axis).

## HORIZONTAL

Refer to Figure 5 for location of items 17 through 22.

- ① **DELAY TIME**-Two controls are used in conjunction with INTENS and DLY'D HORIZONTAL MODE to select the amount of delay time between the start of the sweep and the beginning of the intensified zone.

**Range Selector Switch**-This three-position switch selects 0.5  $\mu$ s, 10  $\mu$ s, and 0.2 ms of delay time. To increase the sweep delay from the calibrated setting of the Range Selector switch, rotate the MULTIPLIER control clockwise.

**MULTIPLIER Control**-Provides variable sweep delay from less than 1 to greater than 20 times the setting of the Range Selector switch.

- ② **SEC/DIV Switch**-Used to select the sweep speed for the sweep generator in a 1-2-5 sequence. For calibrated sweep speeds, the SEC/DIV Variable control must be in the calibrated detent (fully clockwise).

- ③ **SEC/DIV Variable Control**-Provides continuously variable, uncalibrated sweep speeds to at least 2.5 times the calibrated setting. It extends the slowest sweep speed to at least 1.25 s per division.

- ④ **X10 Magnifier Switch**-To increase displayed sweep speed by a factor of 10, pull out the SEC/DIV Variable knob. The fastest sweep speed can be extended to 5 ns per division. Push in the SEC/DIV Variable control knob to regain the X1 sweep speed.


- ⑤ **HORIZONTAL MODE Switch**-This three-position switch determines the mode of operation for the horizontal deflection system.

**NO DLY**-Horizontal deflection is provided by the sweep generator, without a delayed start, at a sweep speed determined by the SEC/DIV switch.

**INTENS**-Horizontal deflection is provided by the sweep generator at a sweep speed determined by the SEC/DIV switch. The sweep generator also provides an intensified zone on the display. The start of the intensified zone represents the sweep-start point when DLY'D HORIZONTAL MODE is selected.

**DLY'D**-Horizontal deflection is provided by the sweep generator at a sweep speed determined by the SEC/DIV switch setting. The start of the sweep is delayed from the initial sweep-trigger point by a time determined by the setting of the DELAY TIME Range Selector switch and MULTIPLIER control.

- ⑥ **POSITION Control**-Positions the display horizontally in all modes.



3397-06

Figure 5. Horizontal controls.

## TRIGGER

Refer to Figure 6 for locations of items 23 through 30.

- ⑦ **EXT INPUT Connector**-Provides a means of introducing external signals into the trigger generator.

- ⑧ **EXT COUPLING Switch**-Determines the method used to couple external signals to the Trigger circuit.

**AC**-Signals above 60 Hz are capacitively coupled to the input of the Trigger circuit. Any dc components are blocked, and signals below 60 Hz are attenuated.

DC-All components of the signal are coupled to the trigger circuitry. This position is useful for displaying low-frequency or low-repetition-rate signals.

**DC÷10**—External trigger signals are attenuated by a factor of 10.

- ② **SOURCE** Switch-Determines the source of the trigger signal that is coupled to the input of the trigger circuit.

**INT**-Permits triggering on signals that are applied to the CH 1 OR X and CH 2 OR Y input connectors. The source of the internal signal is selected by the INT switch.

**LINE**-Provides a triggering signal from a sample of the ac-power-source waveform. This trigger source is useful when channel-input signals are time related (multiple or submultiple) to the frequency on the power-source-input voltage.

**EXT**-Permits triggering on signals applied to the EXT INPUT connector.

- ② **INT** Switch-Selects the source of the triggering signal when the SOURCE switch is set to INT.

**CH 1**-The signal applied to the CH 1 OR X input connector is the source of the trigger signal.

**VERT MODE**-The internal trigger source is determined by the signals selected for display by the VERTICAL MODE switches.

**CH 2**-The signal applied to the CH 2 OR Y input connector is the source of the trigger signal.

- ② **LEVEL** Control-Selects the amplitude point on the trigger signal at which the sweep is triggered.

- ② **TRIG'D** Indicator-The light-emitting diode (LED) illuminates to indicate that the sweep is triggered.

- ② **SLOPE** Switch-Selects the slope of the signal that triggers the sweep (also refer to TV Signal Displays at the end of "Instrument Familiarization").

**I**-Sweep is triggered on the positive-going portion of the trigger signal.

**∩**-Sweep is triggered on the negative-going portion of the trigger signal.

- ③ **MODE** Switch-Determines the trigger mode for the sweep.

**AUTO**-Permits triggering on waveforms having repetition rates of at least 20 Hz. Sweep free-runs in the absence of an adequate trigger signal or when the repetition rate is below 20 Hz. The range of the TRIGGER LEVEL control will compensate for the amplitude variations of the trigger signals.

**NORM**-Sweep is initiated when an adequate trigger signal is applied. In the absence of a trigger signal, no baseline trace will be present. Triggering on television lines is accomplished in this mode.

**TV FIELD**-Permits triggering on television field-rate signals (refer to TV Signal Displays at the end of "Instrument Familiarization").

- ③ **VAR HOLDOFF** Control-Provides continuous control of holdoff time between sweeps. Increases the holdoff time by at least a factor of four. This control improves the ability to trigger on aperiodic signals (such as complex digital waveforms).




Figure 6. Trigger controls, connector, and indicator.

## REAR PANEL

Refer to Figure 7 for location of item 32.

- 32 EXT Z AXIS Connector-Provides a means of connecting external signals to the Z-axis amplifier to

intensity modulate the crt display. Applied signals do not affect display waveshape. Signals with fast rise times and fall times provide the most abrupt intensity change, and a 5-V p-p signal will produce noticeable modulation. The Z-axis signals must be time-related to the display to obtain a stable presentation on the crt.




Figure 7. Rear-panel connector.

# OPERATING CONSIDERATIONS

This section contains basic operating information and techniques that should be considered before attempting any measurements.

## GRATICULE

The graticule is internally marked on the faceplate of the crt to enable accurate measurements without parallax error (see Figure 8). It is marked with eight vertical and ten horizontal major divisions. Each major division is divided into five subdivisions. The vertical deflection factors and horizontal timing are calibrated to the graticule so that accurate measurements can be made directly from the crt. Also, percentage markers for the measurement of rise and fall times are located on the left side of the graticule.




Figure 8. Graticule measurement markings.

## GROUNDING

The most reliable signal measurements are made when the 2213 and the unit under test are connected by a common reference (ground lead), in addition to the signal lead or probe. The probe's ground lead provides the best grounding method for signal interconnection and ensures the maximum amount of signal-lead shielding in the probe cable. A separate ground lead can also be connected from the unit under test to the oscilloscope GND connector located on the front panel.

## SIGNAL CONNECTIONS

Generally, probes offer the most convenient means of connecting an input signal to the instrument. They are shielded to prevent pickup of electromagnetic interference, and the supplied 10X probe offers a high input impedance that minimizes circuit loading. This allows the circuit under test to operate with a minimum of change from its normal condition as measurements are being made.

Coaxial cables may also be used to connect signals to the input connectors, but they may have considerable effect on the accuracy of a displayed waveform. To maintain the original frequency characteristics of an applied signal, only high-quality, low-loss coaxial cables should be used. Coaxial cables should be terminated at both ends in their characteristic impedance. If this is not possible, use suitable impedance-matching devices.

## INPUT COUPLING CAPACITOR PRECHARGING

When the input coupling switch is set to GND, the input signal is connected to ground through the input coupling capacitor in series with a 1-M $\Omega$  resistor to form a precharging network. This network allows the input coupling capacitor to charge to the average dc-voltage level of the signal applied to the probe. Thus, any large voltage transients that may accidentally be generated will not be applied to the amplifier input when the input coupling switch is moved from GND to AC. The precharging network also provides a measure of protection to the external circuitry by reducing the current levels that can be drawn from the external circuitry during capacitor charging.

The following procedure should be used whenever the probe tip is connected to a signal source having a different dc level than that previously applied, especially if the dc-level difference is more than 10 times the VOLTS/DIV switch setting:

1. Set the AC-GND-DC switch to GND before connecting the probe tip to a signal source.

2. Insert the probe tip into the oscilloscope GND connector.

3. Wait several seconds for the input coupling capacitor to discharge.

4. Connect the probe tip to the signal source.

5. Wait several seconds for the input coupling capacitor to charge.

6. Set the AC-GND-DC switch to AC. The display will remain on the screen, and the ac component of the signal can be measured in the normal manner.

## INSTRUMENT COOLING

To maintain adequate instrument cooling, the ventilation holes on both sides and rear panel of the equipment cabinet must remain free of obstructions.

# INSTRUMENT FAMILIARIZATION

## INTRODUCTION

The procedures in this section are designed to assist you in quickly becoming familiar with the 2213. They provide information which demonstrates the use of all the controls, connectors, and indicators and will enable you to efficiently operate the instrument.

Before proceeding with these instructions, verify that the POWER switch is OFF (push button out), then plug the power cord into the ac-power-input-source outlet.

If during the performance of these procedures an improper indication or instrument malfunction is noted, first verify correct operation of associated equipment. Should the malfunction persist, refer the instrument to qualified service personnel for repair or adjustment.

The equipment listed in Table 1, or equivalent equipment, is required to complete these familiarization procedures.

Table 1  
Equipment Required for Instrument Familiarization Procedure

Description	Minimum Specification
Calibration Generator	Standard-amplitude accuracy: $\pm 0.25\%$ . Signal amplitude: 2 mV to 50 V. Output signal: 1 -kHz square wave. Fast-rise repetition rate: 1 to 100 kHz. Rise time: 1 ns or less. Signal amplitude: 100 mV to 1 V. Aberrations: $\pm 2\%$ .
Dual-Input Coupler	Connectors: bnc-female-to-dual-bnc-male.
Cable (2 required)	Impedance: 50 $\Omega$ . Length: 42 in. Connectors: bnc.
Adapter	Connectors: bnc-female-to-bnc female.
Termination	Impedance: 50 $\Omega$ . Connectors: bnc.

## BASELINE TRACE

First obtain a baseline trace, using the following procedure.

1. Preset the instrument front-panel controls as follows:

### Display

AUTO INTENSITY	Fully counterclockwise (minimum)
AUTO FOCUS	Midrange

### Vertical (Both Channels)

AC-GND-DC	<b>AC</b>
VOLTS/DIV	50m (1X)
VOLTS/D IV Variable	Calibrated detent (fully clockwise)
VERTICAL MODE	CH 1
INVERT	Off (button out)
POSITION	Midrange

### Horizontal

SEC/DIV	<b>0.5 ms</b>
SEC/DIV Variable	Calibrated detent (fully clockwise)
HORIZONTAL MODE	NO DLY
X 10 Magnifier	Off (variable knob in)
<b>POSITION</b>	Midrange
<b>DELAY TIME</b>	
Range Selector	0.2 ms
<b>MULTIPLIER</b>	Fully counterclockwise

### Trigger

SLOPE	$\int$ (lever up)
LEVEL	Midrange
MODE	AUTO
EXT COUPLING	AC
SOURCE	INT
VAR HOLDOFF	Fully counterclockwise
INT	VERT MODE

2. Press in the POWER switch button (ON) and allow the instrument to warm up for 20 minutes.

3. Adjust the AUTO INTENSITY control for desired display brightness.

4. Adjust the Vertical and Horizontal POSITION controls to center the trace on the screen.

## NOTE

*Normally, the resulting trace will be parallel with the center horizontal graticule line and should not require adjustment. If trace alignment is required, see the "Trace Rotation" adjustment procedure under "Operator's Adjustments."*

## DISPLAYING A SIGNAL

After obtaining a baseline trace, you are now ready to connect an input signal and display it on the crt screen.

1. Connect the calibration generator standard-amplitude output to both the CH 1 and CH 2 inputs as shown in Figure 9.

2. Set the calibration generator for a standard-amplitude 1-kHz square-wave signal and adjust its output to obtain a vertical display of 4 divisions.

3. Adjust the Channel 1 POSITION control to center the display vertically on the screen.

4. Adjust the TRIGGER LEVEL control, if necessary, to obtain a stable triggered display.

## NOTE

*The TRIG'D indicator should illuminate to indicate that the sweep is triggered.*

5. Rotate the AUTO FOCUS control between its maximum clockwise and counterclockwise positions. The display should become blurred on either side of the optimum control setting.

6. Set the AUTO FOCUS control for a sharp, well-defined display over the entire trace length.

7. Move the display off the screen using the Channel 1 POSITION control.

8. Press in and hold the BEAM FIND push button; the display should reappear on the screen. Adjust the Channel 1 and Horizontal POSITION controls to center the trace both vertically and horizontally. Release the

**BEAM FIND** button; the display should remain within the viewing area.

9. Adjust the **AUTO INTENSITY** control counterclockwise until the display disappears.

10. Press in and hold the **BEAM FIND** push button; the display should reappear. Release the **BEAM FIND** button and adjust the **AUTO INTENSITY** control to desired display brightness.

**Using the Vertical Section**

1. Set the Channel 1 **AC-GND-DC** switch to **GND**.
2. Adjust the trace to the center horizontal graticule line.
3. Set the Channel 1 **AC-GND-DC** switch to **DC**.
4. Observe that the bottom of the display remains at the center horizontal graticule line (ground reference).

5. Set the Channel 1 **AC-GND-DC** switch to **AC**.
6. Observe that the display is centered approximately at the center horizontal line.
7. Set the CH 1 **VOLTS/DIV** switch to 0.1 (IX) and observe that a 2-division vertical display appears.
8. Rotate the CH 1 **VOLTS/DIV** Variable control fully counterclockwise.
9. Observe that minimum vertical deflection occurs when the **VOLTS/DIV** Variable control is fully counterclockwise.
10. Rotate the CH 1 **VOLTS/DIV** Variable control fully clockwise to the **CAL** detent.
11. Select **CH 2 VERTICAL MODE** and again perform preceding steps 1 through 10 using Channel 2 controls. Performance should be similar to Channel 1.




Figure 9. Initial setup for instrument familiarization procedure.

12. Set both Channel 1 and Channel 2 AC-GND-DC switches to DC. Ensure that both CH 1 and CH 2 VOLTS/DIV switches are set to 0.1 (1 X) for 2-division displays.

13. Select BOTH and ADD VERTICAL MODE and observe that the resulting display is 4 divisions in amplitude. Both Channel 1 and Channel 2 POSITION controls should move the display. Recenter the display on the screen.

14. Press in the Channel 2 INVERT push button to invert the Channel 2 signal.

15. Observe that the display is a straight line, indicating that the algebraic sum of the two signals is zero.

16. Set the CH 2 VOLTS/DIV switch to 50 m (1X).

17. Observe the 2-division display, indicating that the algebraic sum of the two signals is no longer zero.

18. Press in the Channel 2 INVERT push button again to release it. Observe a noninverting display having a 6-division signal amplitude.

19. Set both Channel 1 and Channel 2 AC-GND-DC switches to GND.

20. Set the CH 1 VOLTS/DIV switch to 50 m (1X).

21. Select ALT VERTICAL MODE. Position the Channel 1 trace two divisions above the center graticule line and position the Channel 2 trace two divisions below the center graticule line.

22. Rotate the SEC/DIV switch throughout its range (except X-Y). The display will alternate between channels at all sweep speeds. This mode is most useful for sweep speeds from 0.05  $\mu$ s to 0.2 ms per division.

23. Select CHOP VERTICAL MODE and rotate the SEC/DIV switch throughout its range (except X-Y). A dual-trace display will be presented at all sweep speeds, but unlike the ALT mode, both Channel 1 and Channel 2 signals are displayed for each sweep speed on a time-shared basis. This mode is most useful for sweep speeds from 0.5 ms to 0.5 s per division.

24. Select CH 1 VERTICAL MODE and set Channel 1 AC-GND-DC switch to DC. Recenter the display on the screen.

### Using the Horizontal Section

1. Return the SEC/DIV switch to 0.5 ms and note the display for future comparison in step 3.

2. Set the SEC/DIV switch to 5 ms and pull the SEC/DIV Variable control knob out to obtain X10 sweep magnification.

3. Observe that the display is similar to that obtained in step 1.

4. Rotate the Horizontal POSITION control throughout its range. Observe that the display can be positioned to either side of the center vertical graticule line.

5. Push in the SEC/DIV Variable control knob to obtain a X1 sweep.

6. Return the SEC/DIV switch to 0.5 ms.

7. Rotate the VAR HOLDOFF control to its maximum clockwise position.

8. Observe that the crt trace starts to flicker as the holdoff between sweeps is increased.

9. Return the VAR HOLDOFF control to its NORM position (fully counterclockwise). Note the display for future comparison in step 11.

10. Rotate the SEC/DIV Variable control out of the CAL detent to its maximum counterclockwise position.

11. Observe that the sweep rate is approximately 2.5 times slower than in step 9, as indicated by more cycles displayed on the screen.

12. Return the SEC/DIV Variable control to the CAL detent (fully clockwise).

### Using the Delay Time Controls

1. Select **INTENS HORIZONTAL MODE**.
2. Rotate the **MULTIPLIER** control; observe that the start of the intensified zone moves along the display.
3. Select **DLY'D HORIZONTAL MODE** and observe that the intensified zone, previously viewed with **INTENS** selected, is now displayed on the crt screen.
4. Observe that the display moves continuously across the screen as the **MULTIPLIER** control is rotated.
5. Set the **SEC/DIV** switch to  $5 \mu s$  and observe that the magnification of the display is approximately 100 times greater.
6. Select **NO DLY HORIZONTAL MODE** and return the **SEC/DIV** switch to **0.5 ms**.

### Using the Trigger Section

1. Rotate the **TRIGGER LEVEL** control between its maximum clockwise and counterclockwise positions. The display will remain triggered throughout the rotation of the control.
2. Return the **TRIGGER LEVEL** control to the midrange position.
3. Set the **TRIGGER SLOPE** switch to  $\searrow$  (minus). Observe that the display starts on the negative-going slope of the applied signal.
4. Return the **TRIGGER SLOPE** switch to  $\nearrow$  (plus). Observe that the display starts on the positive-going slope of the applied signal.
5. Set the **INT** switch to **CH 1**, select **CH 2 VERTICAL MODE**, and set the Channel 1 **AC-GND-DC** switch to **GND**. Observe that the display free-runs. Return the Channel 1 **AC-GND-DC** switch to **AC**.
6. Set the **INT** switch to **CH 2**, select **CH 1 VERTICAL MODE**, and set the Channel 2 **AC-GND-DC** switch to **GND**. Observe that the display free-runs. Return the Channel 2 **AC-GND-DC** switch to **AC** and set the **INT** switch to **VERT MODE**.

7. Set the **TRIGGER MODE** switch to **NORM**.

8. Rotate the **TRIGGER LEVEL** control between its maximum clockwise and counterclockwise positions. Observe that the **TRIG'D** indicator illuminates only when the display is correctly triggered.

9. Set the **TRIGGER MODE** switch to **AUTO** and set the **TRIGGER SOURCE** switch to **EXT**.

10. Remove the calibration signal from the **CH 2** input connector and connect it to the **EXT INPUT** connector.

11. Set the **CH 1 VOLTS/DIV** switch to **0.5 (IX)** and adjust the output of the calibration generator to provide a 4-division display. Adjust the **TRIGGER LEVEL** control for a stable display and note the range over which a stable display can be obtained (for comparison in step 13).

12. Set the **TRIGGER SOURCE** switch to **EXT $\div$ 10**.

13. Observe that adjustment of the **TRIGGER LEVEL** control provides a triggered display over a narrower range than in preceding step 11, indicating trigger-signal attenuation.

14. Remove the calibration signal from the **EXT INPUT** connector and reconnect it to the **CH 2** input connector. Set the **TRIGGER SOURCE** switch to **INT** and adjust the **TRIGGER LEVEL** control for a stable display.

### Using the X-Y Mode

1. Set both the **CH 1** and **CH 2 VOLTS/DIV** switches to **1 (IX)** and adjust the generator output to provide a 5-division display.
2. Select **X-Y** mode by switching the **SEC/DIV** switch to its fully counterclockwise position.
3. Adjust the **AUTO INTENSITY** control for desired display brightness. Observe that two dots are displayed diagonally. This display can then be positioned horizontally with the **Horizontal POSITION** control and vertically with the **Channel 2 POSITION** control. Note that the dots are separated by 5 horizontal divisions and 5 vertical divisions.
4. Set both the **CH 1** and **CH 2 VOLTS/DIV** switches to **2 (1X)**. Note that the dots are now separated by 2.5 horizontal divisions and 2.5 vertical divisions.
5. Return the **SEC/DIV** switch to **0.5 ms** and adjust the **AUTO INTENSITY** control for desired display brightness.

## Using the Z-Axis Input

1. Disconnect the dual-input coupler from the CH 2 input connector and connect a bnc-female-to-bnc-female adapter to the disconnected end of the coupler.

2. Connect a 42-inch,  $50\text{-}\Omega$  bnc cable from the Z-AXIS INPUT connector (located on the rear panel) to the dual-input coupler via the bnc-female-to-bnc-female adapter.

3. Set the Channel 1 VOLTS/DIV switch to 1 (IX) and adjust the output of the calibration generator to provide a 5-division display.

4. Observe that the positive peaks of the waveform are blanked, indicating intensity modulation (adjust AUTO INTENSITY control as necessary).

5. Disconnect the  $50\text{-}\Omega$  cable from the Z-AXIS INPUT connector and disconnect the dual-input coupler from the CH 1 input connector.

## TV SIGNAL DISPLAYS

### Displaying a TV Line-rate Signal

1. Perform the steps and set the controls as outlined under Baseline Trace and Signal Display to obtain a basic display of the desired TV signal.

2. Set A SEC/DIV to  $10\ \mu\text{s}$ , and A & B INT to CH 1 or CH 2 as appropriate for applied signal.

3. Set A TRIGGER SLOPE for a positive-going signal (lever up) if the applied TV signal sync pulses are positive-going, or for a negative-going signal (lever down) if the TV sync pulses are negative-going.

4. Adjust the A TRIGGER LEVER control for a stable display, and AUTO INTENSITY for desired display brightness. If necessary, adjust VERTICAL VOLTS/DIV control to obtain 5 divisions or greater amplitude for a stable display.

### Displaying a TV Field-rate Signal

1. Perform Step 1 under Displaying a TV Line-rate Signal.

2. Set A SEC/DIV to 2 ms, A TRIGGER MODE to TV FIELD and A & B INT to CH 1 or CH 2 as appropriate for the applied signal.

3. Perform Step 3 and 4 under Displaying a TV Line-rate Signal.

4. To display either Field 1 or Field 2 individually at faster sweep rates (displays of less than one full field), set VERTICAL MODE to BOTH and ALT simultaneously. This synchronizes the Channel 1 display to one field and the Channel 2 display to the other field.

To change the field that is displayed, interrupt the triggering by repeatedly setting the AC GND DC switch to GND or disconnecting the signal from the applied signal input until the other field is displayed. To display both fields simultaneously, apply the input signal to both the CH 1 and CH 2 inputs via two probes, two cables, or through a dual-input coupler.

To examine either a TV Field-rate or Line-rate signal in more detail, either the X10 Magnifier or HORIZONTAL MODE functions may be employed as described for other signals elsewhere in this manual.

# OPERATOR'S ADJUSTMENTS

## INTRODUCTION

Two adjustments should be performed before making measurements with your oscilloscope: Trace Rotation and Probe Compensation. Before proceeding with the following adjustment instructions, verify that the correct line fuse is installed (refer to the "Preparation for Use" information). Verify that the POWER switch is OFF (button out), then plug the power cord into the ac-power-input source. Push in the POWER switch (ON) and allow a 20-minute warm-up time before starting these adjustments.

## TRACE ROTATION

1. Preset instrument controls and obtain a baseline trace (refer to "Instrument Familiarization").

2. Use the Channel 1 POSITION control to move the baseline trace to the center horizontal graticule line.

### NOTE

**Normally, the resulting trace will be parallel to the center horizontal graticule line, and the Trace Rotation adjustment should not be required.**

3. If the resulting trace is not parallel to the center horizontal graticule line, use a small flat-bit screwdriver to adjust the TRACE ROTATION control and align the trace with the center horizontal graticule line.

## PROBE COMPENSATION

Misadjustment of probe compensation is one of the sources of measurement error. Most attenuator probes are equipped with compensation adjustment. To ensure optimum measurement accuracy, always compensate the oscilloscope probe before making measurements. Probe compensation is accomplished as follows:

## 2213 Operators

1. Preset instrument controls and obtain a baseline trace (refer to "Instrument Familiarization").

2. Connect the two 10X probes (supplied with the instrument) to the CH 1 and CH 2 input connectors.

3. Set both VOLTS/DIV switches to 0.1 (10X PROBE) and set both AC-GND-DC switches to DC.

4. Select CH 1 VERTICAL MODE and insert the tip of the Channel 1 probe in the PROBE ADJUST output jack.

5. Using the approximately 1-kHz PROBE ADJUST square-wave signal, obtain a display of the signal (refer to "Instrument Familiarization").

6. Set the SEC/DIV switch to display several cycles of the PROBE ADJUST signal. Use the Channel 1 POSITION control to vertically center the display.

7. Check the waveform presentation for overshoot and rolloff (see Figure 10). If necessary, adjust the probe compensation for flat tops on the waveforms. Refer to the instructions supplied with the probe for details of compensation adjustment.




Figure 10. Probe compensation.

8. Select CH 2 VERTICAL MODE and connect the Channel 2 probe tip to the PROBE ADJUST output jack.

9. Use the Channel 2 POSITION control to vertically center the display and repeat step 7 for the Channel 2 probe.

## BASIC APPLICATIONS

After becoming familiar with all the capabilities of the 2213 Oscilloscope, the operator can then adopt a convenient method for making a particular measurement. The following information describes the recommended procedures and techniques for making basic measurements with your instrument. When a procedure first calls for presetting instrument controls and obtaining a baseline trace, refer to the "Instrument Familiarization" section and perform steps 1 through 4 under "Baseline Trace."

### NONDELAYED MEASUREMENTS

#### AC Peak-to-Peak Voltage

To perform a peak-to-peak voltage measurement, use the following procedure:

#### NOTE

*This procedure may also be used to make voltage measurements between any two points on the waveform.*

1. Preset instrument controls and obtain a baseline trace.

2. Apply the ac signal to either vertical-channel input connector and set the VERTICAL MODE switch to display the channel used.

3. Set the appropriate VOLTS/DIV switch to display about five divisions of the waveform, ensuring that the VOLTS/DIV Variable control is in the CAL detent.

4. Adjust the TRIGGER LEVEL control to obtain a stable display.

5. Set the SEC/DIV switch to a position that displays several cycles of the waveform.

6. Vertically position the display so that the negative peak of the waveform coincides with one of the horizontal graticule lines (see Figure 11, Point A).

7. Horizontally position the display so that one of the positive peaks coincides with the center vertical graticule line (see Figure 11, Point B).

8. Measure the vertical deflection from peak to peak (see Figure 11, Point A to Point B).




Figure 11. Peak-to-peak waveform voltage.

#### NOTE

*If the amplitude measurement is critical or if the trace is thick (as a result of hum or noise on the signal), a more accurate value can be obtained by measuring from the top of a peak to the top of a valley. This will eliminate trace thickness from the measurement.*

9. Calculate the peak-to-peak voltage, using the following formula:

$$\text{Volts (p-p)} = \text{vertical deflection (divisions)} \times \text{VOLTS/DIV switch setting} \times \text{probe attenuation factor}$$

**EXAMPLE:** The measured peak-to-peak vertical deflection is 4.6 divisions (see Figure 11) with a VOLTS/DIV switch setting of 0.5, using a 10X probe.

Substituting the given values:

$$\text{Volts (p-p)} = 4.6 \text{ div} \times 0.5 \text{ V/div} \times 10 = 23 \text{ V.}$$

### Instantaneous DC Voltage

To measure the dc level at a given point on a waveform, use the following procedure:

1. Preset instrument controls and obtain a baseline trace.

2. Apply the signal to either vertical-channel input connector and set the VERTICAL MODE switch to display the channel used.

3. Verify that the VOLTS/DIV Variable control is in the CAL detent and set the AC-GND-DC switch to GND.

4. Vertically position the baseline trace to the center horizontal graticule line.

5. Set the AC-GND-DC switch to DC. If the waveform moves above the centerline of the crt, the voltage is positive. If the waveform moves below the centerline of the crt, the voltage is negative.

#### NOTE

*If using Channel 2, ensure that the Channel 2 INVERT switch is in its noninverting mode (push button out).*

6. Set the AC-GND-DC switch to GND and position the baseline trace to a convenient reference line, using the Vertical POSITION control. For example, if the voltage to be measured is positive, position the baseline trace to the bottom graticule line. If a negative voltage is to be measured, position the baseline trace to the top graticule line. Do not move the Vertical POSITION control after this reference line has been established. The ground reference line can be checked at any later time by switching the AC-GND-DC switch to GND.

7. Set the AC-GND-DC switch to DC.

8. If the voltage-level measurement is to be made with respect to a voltage level other than ground, apply the reference voltage to the unused vertical-channel input connector. Then position its trace to the reference line.

9. Adjust the TRIGGER LEVEL control to obtain a stable display.

10. Set the SEC/DIV switch to a position that displays several cycles of the signal.

11. Measure the divisions of vertical deflection between the reference line and the desired point on the waveform at which the dc level is to be determined (see Figure 12).




Figure 12. Instantaneous voltage measurement.

12. Calculate the instantaneous voltage, using the following formula:

$$\text{Instantaneous Voltage} = \text{vertical deflection (divisions)} \times \text{polarity (+ or -)}$$

$$\times \text{VOLTS/DIV switch setting} \times \text{probe attenuation factor}$$

**EXAMPLE:** The measured vertical deflection from the reference line is 4.6 divisions (see Figure 12), the waveform is above the reference line, the VOLTS/DIV switch is set to 2, and a 10X attenuator probe is being used.

Substituting the given values:

$$\text{Instantaneous Voltage} = 4.6 \text{ div} \times (+1) \times 2 \text{ V/div} \times 10 = 92 \text{ V.}$$

### Algebraic Addition

With the VERTICAL MODE switch set to BOTH and ADD, the waveform displayed is the algebraic sum of the signals applied to the Channel 1 and Channel 2 inputs (CH 1 + CH 2). If the Channel 2 INVERT push button is pressed in, the waveform displayed is the difference between the signals applied to the Channel 1 and Channel 2 inputs (CH 1 - CH 2). The total deflection factor in the ADD mode is equal to the deflection factor indicated by either VOLTS/DIV switch (when both VOLTS/DIV switches are set to the same deflection factor). A common use for the ADD mode is to provide a dc offset for a signal riding on top of a high dc level.

The following general precautions should be observed when using the ADD mode.

- a. Do not exceed the input voltage rating of the oscilloscope.
- b. Do not apply signals that exceed the equivalent of about eight times the VOLTS/DIV switch settings, since large voltages may distort the display. For example, with a VOLTS/DIV switch setting of .5, the voltage applied to that channel should not exceed approximately 4 volts.
- c. Use Channel 1 and Channel 2 POSITION control settings which most nearly position the signal on each channel to midscreen, when viewed in either CH 1 or CH 2 VERTICAL MODE. This ensures the greatest dynamic range for ADD mode operation.
- d. To attain similar response from each channel, set both the Channel 1 and Channel 2 AC-GND-DC switches to the same position.

**EXAMPLE:** Using the graticule center line as 0 V, the Channel 1 signal is at a 3-division, positive dc level (see Figure 13A).

1. Multiply 3 divisions by the VOLTS/DIV switch setting to determine the dc-level value.
2. To the Channel 2 input connector, apply a negative dc level (or positive level, using the Channel 2 INVERT switch) whose value was determined in step 1 (see Figure 13B).
3. Select ADD and BOTH VERTICAL MODE to place the resultant display within the operating range of the vertical POSITION controls (see Figure 13C).

### Common-Mode Rejection

The ADD mode can also be used to display signals that contain undesirable frequency components. The undesirable components can be eliminated through common-mode rejection. The precautions given under the preceding "Algebraic Addition" procedure should be observed.

**EXAMPLE:** The signal applied to the Channel 1 input connector contains unwanted ac-input-power-source frequency components (see Figure 14A). To remove the undesired components, use the following procedure:

1. Preset instrument controls and obtain a baseline trace.




Figure 13. Algebraic addition.

2. Apply the signal containing the unwanted line-frequency components to the Channel 1 input.
3. Apply a line-frequency signal to the Channel 2 input.
4. Select BOTH and ALT VERTICAL MODE and press in the Channel 2 INVERT push button.

4. Set the SEC/DIV switch to display one complete period of the waveform. Ensure that the SEC/DIV Variable control is in the CAL detent.

5. Adjust the Channel 2 VOLTS/DIV switch and Variable control so that the Channel 2 display is approximately the same amplitude as the undesired portion of the Channel 1 display (see Figure 14A).

5. Position the display to place the time-measurement points on the center horizontal graticule line (see Figure 15).

6. Select ADD VERTICAL MODE and slightly readjust the Channel 2 VOLTS/DIV Variable control for maximum cancellation of the undesired signal component (see Figure 14B).




Figure 14. Common-mode rejection.

**Time Duration**

To measure time between two points on a waveform, use the following procedure:

1. Preset instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical-channel input connector and set the VERTICAL MODE switch to display the channel used.
3. Adjust the TRIGGER LEVEL control to obtain a stable display.




Figure 15. Time duration.

6. Measure the horizontal distance between the time-measurement points.

7. Calculate time duration, using the following formula:

$$\text{Time Duration} = \frac{\text{horizontal distance (divisions)} \times \text{SEC/DIV switch setting}}{\text{magnification factor}}$$

**EXAMPLE:** The distance between the time-measurement points is 8.3 divisions (see Figure 15), and the SEC/DIV switch is set to 2 ins. The X10 Magnifier switch is pushed in (1 X magnification).

Substituting the given values:

$$\text{Time Duration} = 8.3 \text{ div} \times 2 \text{ ms/div} = 16.6 \text{ ms}$$

### Frequency

The frequency of a recurrent signal can be determined from its time-duration measurement as follows:

1. Measure the time duration of one waveform cycle using the preceding "Time Duration" measurement procedure.

2. Calculate the reciprocal of the time-duration value to determine the frequency of the waveform.

**EXAMPLE:** The signal in Figure 15 has a time duration of 16.6 ms.

Calculating the reciprocal of time duration:

$$\text{Frequency} = \frac{1}{\text{time duration}} = \frac{1}{16.6 \text{ ms}} = 60 \text{ Hz}$$

### Rise Time

Rise-time measurements use the same methods as time duration, except that the measurements are made between the 10% and 90% points on the leading edge of the waveform (see Figure 16). Fall time is measured between the 90% and 10% points on the trailing edge of the waveform.

1. Preset instrument controls and obtain a baseline trace.

2. Apply an exact 5-division signal to either vertical-channel input connector and set the VERTICAL MODE switch to display the channel used. Ensure that the VOLTS/DIV Variable control is in the CAL detent.

#### NOTE

*For rise time greater than 0.2 μs, the VOLTS/DIV Variable control may be used to obtain an exact S-division display.*

3. Set the TRIGGER SLOPE switch to ⌈ (plus). Use a sweep-speed setting that displays several complete cycles or events (if possible).

4. Adjust vertical positioning so that the zero reference of the waveform touches the 0% graticule line and the top of the waveform touches the 100% graticule line (see Figure 16).




Figure 16. Rise time.

5. Set the SEC/DIV switch for a single-waveform display, with the rise time spread horizontally as much as possible.

6. Horizontally position the display so the 10% point on the waveform intersects the second vertical graticule line (see Figure 16, Point A).

7. Measure the horizontal distance between the 10% and 90% points and calculate the time duration using the following formula:

$$\text{Rise Time} = \frac{\text{horizontal distance (divisions)} \times \text{SEC/DIV switch setting}}{\text{magnification factor}}$$

**EXAMPLE:** The horizontal distance between the 10% and 90% points is 5 divisions (see Figure 16), and the SEC/DIV switch is set to 1  $\mu\text{s}$ . The X10 magnifier knob is pushed in (1 X magnification).

Substituting the given values in the formula:

$$\text{Rise Time} = \frac{5 \text{ div} \times 1 \mu\text{s/div}}{1} = 5 \mu\text{s}$$

### Time Difference Between Two Time-Related Pulses

The calibrated sweep speed and dual-trace features of the 2213 allow measurement of the time difference between two separate events. To measure time difference, use the following procedure:

1. Preset instrument controls and obtain a baseline trace.

2. Set the TRIGGER SOURCE switch to CH 1.

3. Set both AC-GND-DC switches to the same position, depending on the type of input coupling desired.

4. Using either probes or cables with equal time delays, connect a known reference signal to the Channel 1 input and the comparison signal to the Channel 2 input.

5. Set both VOLTS/DIV switches for 4- or 5-division displays.

6. Select BOTH VERTICAL MODE; then select either ALT or CHOP, depending on the frequency of input signals.

7. If the two signals are of opposite polarity, press in the Channel 2 INVERT push button to invert the Channel 2 display (signals may be of opposite polarity due to 180° phase difference; if so, note this for use later in the final calculation).

8. Adjust the TRIGGER LEVEL control for a stable display.

9. Set the SEC/DIV switch to a sweep speed which provides three or more divisions of horizontal separation between the reference points on the two displays. Center each of the displays vertically (see Figure 17).

10. Measure the horizontal difference between the two signal reference points and calculate the time difference using the following formula:

$$\text{Time Difference} = \frac{\text{SEC/DIV switch setting} \times \text{horizontal difference (divisions)}}{\text{magnification factor}}$$

**EXAMPLE:** The SEC/DIV switch is set to 50  $\mu\text{s}$ , the X10 magnifier knob is pulled out, and the horizontal difference between waveform measurement points is 4.5 divisions.

Substituting the given values in the formula:

$$\text{Time Difference} = \frac{50 \mu\text{s/div} \times 4.5 \text{ div}}{10} = 22.5 \mu\text{s}$$




Figure 17. Time difference between two time-related pulses.

**Phase Difference**

In a similar manner to “Time Difference,” phase comparison between two signals of the same frequency can be made using the dual-trace feature of the 2213. This method of phase difference measurement can be used up to the frequency limit of the vertical system. To make a phase comparison, use the following procedure:

1. Preset instrument controls and obtain a baseline trace, then set the TRIGGER SOURCE switch to CH 1.

2. Set both AC-GND-DC switches to the same position, depending on the type of input coupling desired.

3. Using either probes or coaxial cables with equal time delays, connect a known reference signal to the Channel 1 input and the unknown signal to the Channel 2 input.

4. Select BOTH VERTICAL MODE; then select either ALT or CHOP, depending on the frequency of the input signals. The reference signal should precede the comparison signal in time.

5. If the two signals are of opposite polarity, press in the Channel 2 INVERT push button to invert the Channel 2 display.

6. Set both VOLTS/DIV switches and both Variable controls so the displays are equal in amplitude.

7. Adjust the TRIGGER LEVEL control for a stable display.

8. Set the SEC/DIV switch to a sweep speed which displays about one full cycle of the waveforms.

9. Position the displays and adjust the SEC/DIV Variable control so that one reference-signal cycle occupies exactly 8 horizontal graticule divisions at the 50% rise-time points (see Figure 18). Each division of the graticule now represents 45° of the cycle (360° ÷ 8 divisions), and the horizontal graticule calibration can be stated as 45° per division.

10. Measure the horizontal difference between corresponding points on the waveforms at a common horizontal graticule line (50% of rise time) and calculate the phase difference using the following formula:

$$\text{Phase Difference} = \text{horizontal difference (divisions)} \times \text{horizontal graticule calibration (deg/div)}$$

**EXAMPLE:** The horizontal difference is 0.6 division with a graticule calibration of 45° per division as shown in Figure 18.

Substituting the given values into the phase difference formula:

$$\text{Phase Difference} = 0.6 \text{ div} \times 45^\circ/\text{div} = 27^\circ$$

More accurate phase measurements can be made by using the X10 Magnifier function to increase the sweep rate without changing the SEC/DIV Variable control setting.

**EXAMPLE:** If the sweep rate were increased 10 times with the magnifier (X10 Magnifier out), the magnified horizontal graticule calibration would be 45°/division divided by 10 (or 4.5°/division). Figure 19 shows the same signals illustrated in Figure 18, but magnifying the displays results in a horizontal difference of 6 divisions between the two signals.

Substituting the given values in the phase difference formula:

$$\text{Phase Difference} = 6 \text{ div} \times 4.5^\circ/\text{div} = 27^\circ$$




Figure 18. Phase difference.




Figure 19. High-resolution phase difference.

**Amplitude Comparison**

In some applications it may be necessary to establish a set of deflection factors other than those indicated by the VOLTS/DIV switch settings. This is useful for comparing unknown signals to a reference signal of known amplitude. To accomplish this, a reference signal of known amplitude is first set to an exact number of vertical divisions by adjusting the VOLTS/DIV switch and Variable control. Unknown signals can then be quickly and accurately compared with the reference signal without disturbing the setting of the VOLTS/DIV Variable control. The procedure is as follows.

1. Preset instrument controls and obtain a baseline trace.
2. Apply the reference signal to either vertical channel input and set the VERTICAL MODE switch to display the channel used.
3. Set the amplitude of the reference signal to an exact number of vertical divisions by adjusting the VOLTS/DIV switch and VOLTS/DIV Variable control.
4. Establish a vertical conversion factor, using the following formula (reference signal amplitude must be known):

$$\text{Vertical Conversion Factor} = \frac{\text{reference signal amplitude (volts)}}{\text{vertical deflection (divisions)} \times \text{VOLTS/DIV switch setting}}$$

5. Disconnect the reference signal and apply the unknown signal to be measured to the same channel input. Adjust the VOLTS/DIV switch to a setting that provides sufficient vertical deflection to make an accurate measurement. Do not readjust the VOLTS/DIV Variable control.

6. Establish an arbitrary deflection factor, using the following formula:

$$\text{Arbitrary Deflection Factor} = \frac{\text{vertical conversion factor}}{\text{VOLTS/DIV switch setting}}$$

7. Measure the vertical deflection of the unknown signal in divisions and calculate its amplitude using the following formula:

$$\text{Unknown Signal Amplitude} = \frac{\text{arbitrary deflection factor}}{\text{vertical deflection (divisions)}}$$

EXAMPLE: The reference signal amplitude is 30 V, with a VOLTS/DIV switch setting of 5 and the VOLTS/DIV Variable control adjusted to provide a vertical deflection of exactly 4 divisions,

Substituting these values in the vertical conversion factor formula:

$$\text{Vertical Conversion Factor} = \frac{30 \text{ v}}{4 \text{ div} \times 5 \text{ V/div}} = 1.5$$

Continuing, for the unknown signal the VOLTS/DIV switch setting is 1, and the peak-to-peak amplitude spans five vertical divisions. The arbitrary deflection factor is then determined by substituting values in the formula:

$$\text{Arbitrary Deflection Factor} = 1.5 \times 1 \text{ V/div} = 1.5 \text{ V/div}$$

The amplitude of the unknown signal can then be determined by substituting values in the unknown signal amplitude formula:

$$\text{Amplitude} = 1.5 \text{ V/div} \times 5 \text{ div} = 7.5 \text{ V}$$

## Time Comparison

In a similar manner to "Amplitude Comparison," repeated time comparisons between unknown signals and a reference signal (e.g., on assembly line test) may be easily and accurately measured with the 2213. To accomplish this, a reference signal of known time duration is first set to an exact number of horizontal divisions by adjusting the SEC/DIV switch and the SEC/DIV Variable control. Unknown signals can then be compared with the reference signal without disturbing the setting of the SEC/DIV Variable control. The procedure is as follows:

1. Set the time duration of the reference signal to an exact number of horizontal divisions by adjusting the SEC/DIV switch and the SEC/DIV Variable control.

2. Establish a horizontal conversion factor, using the following formula (reference signal time duration must be known):

$$\text{Horizontal Conversion Factor} = \frac{\text{reference signal time duration (seconds)}}{\text{horizontal distance (divisions)} \times \text{SEC/DIV switch setting}}$$

3. For the unknown signal, adjust the SEC/DIV switch to a setting that provides sufficient horizontal deflection to make an accurate measurement. Do not readjust the SEC/DIV Variable control.

4. Establish an arbitrary deflection factor, using the following formula:

$$\text{Arbitrary Deflection Factor} = \text{horizontal conversion factor} \times \text{SEC/DIV switch setting}$$

5. Measure the horizontal distance of the unknown signal in divisions and calculate its time duration using the following formula:

$$\text{Time Duration} = \text{arbitrary deflection factor} \times \text{horizontal distance (divisions)}$$

6. Frequency of the unknown signal can then be determined by calculating the reciprocal of its time duration.

**EXAMPLE:** The reference signal time duration is 2.19 ms, the SEC/DIV switch setting is 0.2 ms, and the

SEC/DIV Variable control is adjusted to provide a horizontal distance of exactly 8 divisions.

Substituting the given values in the horizontal conversion factor formula:

$$\text{Horizontal Conversion Factor} = \frac{2.19 \text{ ms}}{8 \text{ div} \times 0.2 \text{ ms/div}} = 1.37$$

Continuing, for the unknown signal the SEC/DIV switch setting is 50  $\mu$ s, and one complete cycle spans 7 horizontal divisions. The arbitrary deflection factor is then determined by substituting values in the formula:

$$\text{Arbitrary Deflection Factor} = 1.37 \times 50 \mu\text{s/div} = 68.5 \mu\text{s/div}$$

The time duration of the unknown signal can then be computed by substituting values in the formula:

$$\text{Time Duration} = 68.5 \mu\text{s/div} \times 7 \text{ div} = 480 \mu\text{s}$$

The frequency of the unknown signal is then calculated:

$$\text{Frequency} = \frac{1}{480 \mu\text{s}} = 2.083 \text{ kHz}$$

## DELAYED-SWEEP MAGNIFICATION

The delayed-sweep feature of the 2213 can be used to provide higher apparent magnification than is provided by the X10 Magnifier switch. Apparent magnification occurs as a result of displaying a selected portion of the trace (INTENS HORIZONTAL MODE) at a faster sweep speed (DLY'D HORIZONTAL MODE).

When INTENS HORIZONTAL MODE is selected, the intensified zone indicates both the location and the start of the sweep that will be displayed in DLY'D HORIZONTAL MODE. Positioning of the intensified zone (i.e., setting the amount of time between start of the sweep and start of the intensified zone) is accomplished with the MULTIPLIER control and the DELAY TIME Range Selector switch. At higher sweep speeds the delay time can be adjusted to allow the starting point of the intensified zone to occur past the end of the display.

With either **INTENS** or **DLY'D HORIZONTAL MODE** selected, the **DELAY TIME** Range Selector switch and the **MULTIPLIER** control provide continuously variable positioning of the start of the delayed sweep. The **DELAY TIME** Range Selector switch allows the start of the intensified zone to be placed near the point of interest, while the **MULTIPLIER** control provides fine adjustment of the intensified zone.

When viewing aperiodic signals (such as complex digital waveforms) with **DLY'D HORIZONTAL MODE** selected, the start of the sweep may not be at the same point as the start of the intensified zone. It may be necessary to connect a reference signal (of the system under test) to the **EXT INPUT** connector to ensure correct display of the selected portion of the waveform.

Using delayed-sweep magnification may produce a display with some slight horizontal movement (pulse jitter). Pulse jitter includes not only the inherent uncertainty of triggering the delayed sweep at exactly the same trigger point each time, but also jitter that may be present in the input signal. If pulse jitter needs to be measured, use the "Pulse Jitter Time Measurement" procedure which follows the discussion of "Magnified Sweep."

### Magnified Sweep

The following procedure explains how to operate the delayed-sweep feature and to determine the resulting apparent magnification factor.

1. Preset instrument controls and obtain a baseline trace.
2. Apply the signal to either vertical channel input connector and set the **VERTICAL MODE** switch to display the channel used.
3. Set the appropriate **VOLTS/DIV** switch to produce a display of approximately 5 divisions in amplitude and center the display.
4. Set the **SEC/DIV** switch to a sweep speed which displays at least one complete waveform cycle.
5. Select **INTENS HORIZONTAL MODE** and set the **DELAY TIME** Range Selector switch for the appropriate delayed time. Adjust the **MULTIPLIER** control to position the start of the intensified zone to the portion of the display to be magnified (see Figure 20A).

6. Select the **DLY'D HORIZONTAL MODE** and increase the sweep speed to magnify the intensified portion of the sweep (see Figure 20B).

7. The apparent sweep magnification can be calculated from the following formula:

$$\text{Apparent Delayed Sweep Magnification} = \frac{\text{initial SEC/DIV setting}}{\text{second SEC/DIV setting}}$$

**EXAMPLE:** Determine the apparent magnification of a display with an initial **SEC/DIV** switch setting of 0.1 ms and the second **SEC/DIV** switch setting of 1  $\mu\text{s}$ .

Substituting the given values:

$$\text{Apparent Magnification} = \frac{1 \times 10^{-4} \text{ s}}{1 \times 10^{-6} \text{ s}} = 10^2 = 100$$




Figure 20. Delayed-sweep magnification.

**Pulse Jitter Time Measurement**

To measure pulse jitter time:

1. Perform steps 1 through 6 of the preceding "Magnified Sweep" procedure.

2. Referring to Figure 21, measure the difference between Point A and Point B in divisions and calculate the pulse jitter time using the following formula:

$$\text{Pulse Jitter Time} = \frac{\text{horizontal difference (divisions)}}{\text{second switch setting SEC/DIV}}$$




Figure 21. Pulse jitter.

# SPECIFICATION

The following electrical characteristics (Table 2) are valid for the 2213 when it has been adjusted at an ambient temperature between +20°C and +30°C, has had a warm-up period of at least 20 minutes, and is operating at an ambient temperature between 0°C and +50°C (unless otherwise noted).

Item listed in the "Performance Requirements" column are verifiable qualitative or quantitative limits, while items listed in the "Supplemental Information" column are either explanatory notes, calibration setup descriptions,

performance characteristics for which no absolute limits are specified, or characteristics that are impractical to check.

Environmental characteristics are given in Table 3. The 2213 meets the requirements of MI L-T-288008, Class 5 equipment, except where otherwise noted.

Physical characteristics of the instrument are listed in Table 4.

Table 2  
Electrical Characteristics

Characteristics	Performance Requirements	Supplemental Information
<b>VERTICAL DEFLECTION SYSTEM</b>		
Deflection Factor		1X gain adjusted with VOLTS/DIV switch set to 20 mV per division.  10X gain adjusted with VOLTS/DIV switch set to 2 mV per division.
Range	2 mV per division to 10 V per division in a 1-2-5 sequence.	
Accuracy		
+20°C to +30°C	±3%.	
0°C to +50°C	±4%. <sup>a</sup>	
Range of VOLTS/DIV Variable Control.	Continuously variable between settings. Increases deflection factor by at least 2.5 to 1.	
Step Response		Measured with a vertically centered 5-division reference signal from a 50-Ω source driving a 50-Ω coaxial cable that is terminated in 50 Ω at the input connector, with the VOLTS/DIV Variable control in its CAL detent.
Rise Time		5.8 ns or less.  Rise time is calculated from the formula:  Rise Time = $\frac{0.35}{\text{BW (in MHz)}}$
Bandwidth		Measured with a vertically centered 6-division reference signal from a 50-Ω source driving a 50-Ω coaxial cable that is terminated in 50 Ω, both at the input connector and at the P6120 probe input, with the VOLTS/DIV Variable control in its CAL detent.
0°C to +40°C		
20 mV to 10 V per Division	Dc to at least 60 MHz.	
2 mV to 10 mV per Division	Dc to at least 50 MHz.	
+40°C to +50°C		
2 mV to 10 V per Division	Dc to at least 50 MHz. <sup>a</sup>	
Chop Mode Repetition Rate		250 kHz ±30%.

<sup>a</sup>Performance Requirement not checked in Service Manual.

Table 2 (cont)

Characteristics	Performance Requirements	Supplemental Information
<b>VERTICAL DEFLECTION SYSTEM (cont)</b>		
Input Characteristics		
Resistance	1 M $\Omega$ $\pm$ 2%. <sup>a</sup>	
Capacitance	30 pF $\pm$ 3 pF. <sup>a</sup>	
Maximum Safe Input Voltage 		
DC Coupled	400 V (dc + peak ac) or 800 V p-p ac to 1 kHz or less. <sup>a</sup>	
AC Coupled	400 V (dc + peak ac) or 800 V p-p ac to 1 kHz or less. <sup>a</sup>	
Common-Mode Rejection Ratio (CMRR)	At least 10 to 1 at 10 MHz.	Checked at 20 mV per division for common-mode signals of 8 divisions or less, with VOLTS/DIV Variable control adjusted for best CMRR at 50 kHz.
<b>TRIGGER SYSTEM</b>		
Trigger Sensitivity AUTO and NORM	<b>0.4 division internal or 50 mV external to 2 MHz, increasing to 1.5 divisions internal or 250 mV external at 60 MHz.</b>	External trigger signal from a 50- $\Omega$ source driving a 50- $\Omega$ coaxial cable that is terminated in 50 $\Omega$ at the input connector.  Will trigger on tv line sync components in NORM only: $\geq$ 0.4 division internal or 50 mV p-p external.
AUTO Lowest Usable Frequency	20 Hz. <sup>a</sup>	
TV FIELD	2.0 divisions of composite video or composite sync. <sup>a</sup>	
External Input		
Maximum Input Voltage 	400 V (dc + peak ac) or 800 V p-p ac at 1 kHz or less. <sup>a</sup>	
Input Resistance	1 M $\Omega$ $\pm$ 2%. <sup>a</sup>	
Input Capacitance	30 pF $\pm$ 3 pF. <sup>a</sup>	
AC Coupled	10 Hz or less at lower -3 dB point. <sup>a</sup>	
LEVEL Control Range (with NORM TRIGGER MODE)		
INT	On screen limits. <sup>a</sup>	
EXT and DC	At least $\pm$ 2 V (4 V p-p). <sup>a</sup>	

<sup>a</sup>Performance Requirement not checked in Service Manual.

Table 2 (cont)

Characteristics	Performance Requirements	Supplemental Information
<b>TRIGGER SYSTEM (cont)</b>		
LEVEL Control Range (with NORM TRIGGER MODE) (cont) EXT and DC+ 10	At least $\pm 20$ V (40 V p-p). <sup>a</sup>	
VAR HOLDOFF Control Range	Increases sweep holdoff time by at least a factor of four. <sup>a</sup>	
<b>HORIZONTAL DEFLECTION SYSTEM</b>		
Sweep Rate Calibrated Range A Sweep	0.5 s per division to 0.05 $\mu$ s per division in a 1-2-5 sequence. X10 Magnifier extends maximum sweep speed to 5 ns per division.	
Accuracy	<b>Unmagnified</b> <b>Magnified</b>	Sweep accuracy applies over the center 8 divisions. Exclude the first 50 ns of the sweep for both magnified and un-magnified sweep speeds and exclude anything beyond the 100th magnified division.
+20°C to +30°C	$\pm 3\%$ $\pm 5\%$	
0°C to +50°C	$\pm 4\%$ <sup>a</sup> $\pm 6\%$ <sup>a</sup>	
POSITION Control Range	Start of sweep to 100th division will position past the center vertical graticule line with X10 Magnifier.	
Variable Control Range	Continuously variable between calibrated settings. Extends the sweep speeds by at least a factor of 2.5.	
Delay Time Range Selector	Minimum delay is less than selected values of 0.5 $\mu$ s, 10 $\mu$ s, and 0.2 ms.	
MULTIPLIER Control	Increases delay time by at least a factor of 20.	
Jitter	One part, or less, in 5,000 (0.02%) of the maximum available delay time.	
<b>X-Y OPERATION (X10 MAGNIFICATION)</b>		
Deflection Factors Range	Same as Vertical Deflection System, with both VOLTS/DIV Variable controls in CAL detent.	

<sup>a</sup>Performance Requirement not checked in Service Manual.

Table 2 (cont)

Characteristics	Performance Requirements	Supplemental Information
-----------------	--------------------------	--------------------------

**X-Y OPERATION (XI MAGNIFICATION) (cont)**

Deflection Factors (cont)			Measured with a dc-coupled, 5-division reference signal.
Accuracy	<b>X-Axis</b>	<b>Y-Axis</b>	
+20°C to +30°C	±5%	±3%	
0°C to +50°C	±6% <sup>a</sup>	±4% <sup>a</sup>	
Bandwidth			Measured with a 5-division reference signal.
X-Axis	Dc to at least 2 MHz.		
Y-Axis	Same as Vertical Deflection System.		
Phase Difference Between X- and Y-Axis <b>Amplifiers</b>	+3° from dc to 50 kHz. <sup>a</sup>		With dc-coupled inputs.

**PROBE ADJUST**

Signal at PROBE ADJUST Jack		
Voltage	0.5 v ±20%.	
Repetition Rate	1 kHz ±20%. <sup>a</sup>	

**Z-AXIS INPUT**

Sensitivity	5 V causes noticeable modulation. Positive-going input signal decreases intensity.	
Usable Frequency Range	Dc to 5 MHz. <sup>a</sup>	
Maximum Safe Input Voltage	30 V (dc + peak ac) or 30 V p-p ac at 1 kHz or less. <sup>a</sup>	
Input Impedance	10 kΩ ±10%. <sup>a</sup>	

**POWER SOURCE**

Line Voltage Range	90 V to 250 V. <sup>a</sup>	
Line Frequency Range	48 Hz to 62 Hz. <sup>a</sup>	
Maximum Power Consumption	50 W. <sup>a</sup>	
Line Fuse	2 A, 250 V, fast.	

<sup>a</sup>Performance Requirement not checked in Service Manual.

Table 2 (cont)

Characteristics	Performance Requirements	Supplemental Information
<b>CATHODE-RAY TUBE</b>		
Display Area	80 by 100 mm. <sup>a</sup>	
Standard Phosphor	P31. <sup>a</sup>	
Nominal Accelerating Voltage	10,000 V. <sup>a</sup>	

<sup>a</sup>Performance Requirement not checked in Service Manual.

Table 3  
Environmental Characteristics

Characteristics	Description
	<b>NOTE</b> <i>The instrument meets all of the following MIL-T-288006 requirements for Class 5 equipment.</i>
<b>Temperature</b>	
<b>Operating</b>	0°C to +50°C (+32°F to +122°F).
<b>Nonoperating</b>	-55°C to +75°C (-67°F to +167°F).
<b>Altitude</b>	
<b>Operating</b>	To 4,500 m (15,000 ft). Maximum operating temperature decreased 1°C per 300 m (1,000 ft) above 1,500 m (5,000 ft).
<b>Nonoperating</b>	To 15,000 m (50,000 ft).
<b>Humidity (Operating and Nonoperating)</b>	5 cycles (120 hours) referenced to MIL-T-28800B, Class 5 instruments.
<b>Vibration (Operating)</b>	15 minutes along each of 3 major axes at a total displacement of 0.015 inch p-p (2.4 g at 55 Hz), with frequency varied from 10 Hz to 55 Hz to 10 Hz in 1-minute sweeps. Hold for 10 minutes at 55 Hz. All major resonances must be above 55 Hz.
<b>Shock (Operating and Nonoperating)</b>	30 g, half-sine, 1 1-ms duration; 3 shocks per axis each direction, for a total of 18 shocks.

**Table 4**  
**Physical Characteristics**

Characteristics	Description
Weight	
With Front-Panel Cover, Accessories, and Pouch	7.6 kg (16.8 lb).
Without Front-Panel Cover, Accessories, and Pouch	6.1 kg (13.5 lb).
Domestic Shipping	8.2 kg (18.0 lb).
Height With Feet and Handle	137 mm (5.4 in).
Width	
With Handle	361 mm (14.2 in).
Without Handle	328 mm (12.9 in).
Depth	
With Front-Panel Cover	445 mm (17.5 in).
Without Front-Panel Cover	439 mm (17.3 in).
With Handle Extended	511 mm (20.1 in).

## ACCESSORIES

### STANDARD ACCESSORIES INCLUDED

2 Probes, 10X 1.5-m length with accessories. ....	010-6120-01
2 Probe accessories, Grabber Tips .....	013-0191-00
1 Operators Manual .....	070-3397-00
1 Service Manual .....	070-3827-00

### OPTIONAL ACCESSORIES

Protective Front-Panel Cover. ....	200-2520-00
Cord Wrap and Storage Pouch .....	016-0677-00
Protective Front-Panel Cover, Cord Wrap, and Storage Pouch. ....	020-0672-00
Low-Cost, General-Purpose Camera. ....	Order C-5C Option 04
SCOPE-MOB1 LE Cart-Occupies less than 18 inches of aisle space, with storage area in base .....	Order 200C
Rack-Mount Adapter Kit .....	016-0466-00

# **K4XL's** **BAMA**

This manual is provided **FREE OF CHARGE** from the "BoatAnchor Manual Archive" as a service to the Boatanchor community.

It was uploaded by someone who wanted to help you repair and maintain your equipment.

If you paid anyone other than BAMA for this manual, you paid someone who is making a profit from the free labor of others without asking their permission.

You may pass on copies of this manual to anyone who needs it. But do it without charge.

Thousands of files are available without charge from BAMA. Visit us at <http://bama.sbc.edu>